

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	00757	Calib. Date: 09/16/2011
Certification File No.:	00758	Calib. No.: 00017
Linearity File No.:	00759	Cert. Date: 09/16/2011
Solution File No.:	00763	Cert. No.: 00013
Sequential File No.:	00763	Lin. Date: 09/16/2011
		Lin. No.: 00013
		Soln. Date: 10/22/2011
		Soln. No.: 00121
		File Date: 10/22/2011

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUL S3-0183
Control Solution %:	0.100%		Expires: 07/07/2013
Solution Control Lot:	11G093		Bottle No.: 0878

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:45D		
Control 1 EC	0.098%	17:46D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	17:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:47D		
Control 2 EC	0.097%	17:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	17:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:48D		
Control 3 EC	0.097%	17:49D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: GENDRON

First Name: FRANK

MI: G

Signature: \_\_\_\_\_



Badge No.: 1120

Date: 10/22/2011