

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0042
Location: HIGHTSTOWN POLICE
Calibration File No.: 00757 Calib. Date: 09/16/2011 Calib. No.: 00017
Certification File No.: 00758 Cert. Date: 09/16/2011 Cert. No.: 00013
Linearity File No.: 00759 Lin. Date: 09/16/2011 Lin. No.: 00013
Solution File No.: 00765 Soln. Date: 12/05/2011 Soln. No.: 00122
Sequential File No.: 00765 File Date: 12/05/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0183
Control Solution %: 0.100% Expires: 07/07/2013
Solution Control Lot: 11G093 Bottle No.: 0879

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:01S		
Control 1 EC	0.100%	12:01S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:02S		
Control 2 EC	0.097%	12:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:03S		
Control 3 EC	0.098%	12:04S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: NIRO

First Name: SGT

MI: J

Signature: _____



Badge No.: 1114

Date: 12/05/2011