Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTN-0042
Location:	HIGHTSTOWN POLICE				
Calibration File No.:	00942	Calib. Date:	09/13/2013	Calib. No.:	00022
Certification File No.:	00943	Cert. Date:	09/13/2013	Cert. No.:	00017
Linearity File No.:	00944	Lin. Date:	09/13/2013	Lin. No.:	00017
Solution File No.:	00945	Soln. Date:	09/13/2013	Soln. No.:	00158
Sequential File No.:	00945	File Date:	09/13/2013		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0182
Control Solution %:	0.100%			Expires:	09/13/2014
Solution Control Lot:	12I107			Bottle No.:	0874

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:22D		
Control 1 EC	0.099%	11:23D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:23D		
Control 2 EC	0.098%	11:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:25D		
Control 3 EC	0.099%	11:25D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDU5 P2 - 035 TJS

Changed By:

Last Name: SNYDER First Name: THOMAS MI: J.

Signature: Sgt. Thomas J. Snyder #5792 Badge No.: 5792

Date: 09/13/2013