

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	00901	Calib. Date: 03/14/2013
Certification File No.:	00902	Calib. No.: 00021
Linearity File No.:	00903	Cert. Date: 03/14/2013
Solution File No.:	00904	Cert. No.: 00016
Sequential File No.:	00904	Lin. Date: 03/14/2013
		Lin. No.: 00016
		Soln. Date: 03/14/2013
		Soln. No.: 00150
		File Date: 03/14/2013

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUL S3-0183
Control Solution %:	0.100%		Expires: 09/13/2014
Solution Control Lot:	12I107		Bottle No.: 0867

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:55S	13:55 D MPB	
Control 1 EC	0.101%	12:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:56S		
Control 2 EC	0.101%	12:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:58S		
Control 3 EC	0.101%	12:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:59S	13:59 D MPB	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDWJP2 - 055 (MPB)

Changed By:

Last Name: GIBSON

First Name: MICHAEL

MI: P

Signature: *Michael P. Gibson #6353*

Badge No.: 6353

Date: 03/14/2013