

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	00835	Calib. Date: 08/31/2012
Certification File No.:	00836	Calib. No.: 00019
Linearity File No.:	00837	Cert. Date: 08/31/2012
Solution File No.:	00890	Cert. No.: 00015
Sequential File No.:	00890	Lin. Date: 08/31/2012
		Lin. No.: 00015
		Soln. Date: 01/28/2013
		Soln. No.: 00145
		File Date: 01/28/2013

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0183
Control Solution %:	0.100%			Expires:	07/07/2013
Solution Control Lot:	11G093			Bottle No.:	0898

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:59S		
Control 1 EC	-.---%	06:00S	N/A	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	06:00S	N/A	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RUNGE

First Name: GLEN

MI: L

Signature: _____



Badge No.: 1118

Date: 01/28/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	00835	Calib. Date: 08/31/2012
Certification File No.:	00836	Calib. No.: 00019
Linearity File No.:	00837	Cert. Date: 08/31/2012
Solution File No.:	00891	Cert. No.: 00015
Sequential File No.:	00891	Lin. Date: 08/31/2012
		Lin. No.: 00015
		Soln. Date: 01/28/2013
		Soln. No.: 00146
		File Date: 01/28/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	11G093	Expires: 07/07/2013
		Bottle No.: 0897

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:52S		
Control 1 EC	-.---%	06:53S	20.1°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	06:53S	20.1°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: RUNGE

First Name: GLEN

MI: L

Signature: _____



Badge No.: 1118

Date: 01/28/2013

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	00835	Calib. Date: 08/31/2012
Certification File No.:	00836	Calib. No.: 00019
Linearity File No.:	00837	Cert. Date: 08/31/2012
Solution File No.:	00892	Cert. No.: 00015
Sequential File No.:	00892	Lin. Date: 08/31/2012
		Lin. No.: 00015
		Soln. Date: 01/28/2013
		Soln. No.: 00147
		File Date: 01/28/2013

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0183
Control Solution %: 0.100%		Expires: 07/07/2013
Solution Control Lot: 11G093		Bottle No.: 0891

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:23S		
Control 1 EC	0.099%	12:24S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:25S		
Control 2 EC	0.097%	12:25S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:26S		
Control 3 EC	0.098%	12:27S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RUNGE

First Name: GLEN

MI: L

Signature: _____

Badge No.: 1118

Date: 01/28/2013