

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	00901	Calib. Date: 03/14/2013
Certification File No.:	00902	Calib. No.: 00021
Linearity File No.:	00903	Cert. Date: 03/14/2013
Solution File No.:	00913	Cert. No.: 00016
Sequential File No.:	00913	Lin. Date: 03/14/2013
		Lin. No.: 00016
		Soln. Date: 05/10/2013
		Soln. No.: 00154
		File Date: 05/10/2013

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0182
Control Solution %:	0.100%			Expires:	09/13/2014
Solution Control Lot:	12I107			Bottle No.:	0861

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:40D		
Control 1 EC	0.100%	12:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:41D		
Control 2 EC	0.100%	12:42D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:43D		
Control 3 EC	0.100%	12:43D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Temperature Probe Serial Number: DDUJP2 - 035 (MPA)

Changed By:

Last Name: GIBSON

First Name: MICHAEL

MI: P

Signature: 

Badge No.: 6353

Date: 05/10/2013