

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	00901	Calib. Date: 03/14/2013
Certification File No.:	00902	Calib. No.: 00021
Linearity File No.:	00903	Cert. Date: 03/14/2013
Solution File No.:	00919	Cert. No.: 00016
Sequential File No.:	00919	Lin. Date: 03/14/2013
		Lin. No.: 00016
		Soln. Date: 06/10/2013
		Soln. No.: 00155
		File Date: 06/10/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0182
Solution Control Lot:	12I107	Expires: 09/13/2014
		Bottle No.: 0862

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:36D		
Control 1 EC	0.100%	10:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:38D		
Control 2 EC	0.099%	10:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:39D		
Control 3 EC	0.099%	10:40D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NIRO

First Name: SGT

MI: J

Signature: _____



Badge No.: 1114

Date: 06/10/2013