

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0042
Location: HIGHTSTOWN POLICE
Calibration File No.: 01108 Calib. Date: 02/18/2015 Calib. No.: 00026
Certification File No.: 01109 Cert. Date: 02/18/2015 Cert. No.: 00021
Linearity File No.: 01110 Lin. Date: 02/18/2015 Lin. No.: 00021
Solution File No.: 01132 Soln. Date: 06/23/2015 Soln. No.: 00187
Sequential File No.: 01132 File Date: 06/23/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0182
Control Solution %: 0.100% Expires: 06/20/2016
Solution Control Lot: 14F130 Bottle No.: 0062

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:20D		
Control 1 EC	0.099%	16:21D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:21D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:21D		
Control 2 EC	0.098%	16:22D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:22D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:23D		
Control 3 EC	0.098%	16:23D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:23D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:24D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GENDRON

First Name: FRANK

MI: G

Signature: 

Badge No.: 1120

Date: 06/23/2015