

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01134	Calib. Date: 07/13/2015
Certification File No.:	01135	Calib. No.: 00027
Linearity File No.:	01136	Cert. Date: 07/13/2015
Solution File No.:	01155	Cert. No.: 00022
Sequential File No.:	01155	Lin. Date: 07/13/2015
		Lin. No.: 00022
		Soln. Date: 11/13/2015
		Soln. No.: 00192
		File Date: 11/13/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0182
Solution Control Lot:	15F139	Expires: 06/13/2017
		Bottle No.: 1375

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:22S		
Control 1 EC	0.099%	07:23S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:23S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:23S		
Control 2 EC	0.099%	07:24S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:24S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:25S		
Control 3 EC	0.099%	07:25S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:25S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: \_\_\_\_\_

Badge No.: 1113

Date: 11/13/2015