Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTN-0042
Location:	HIGHTSTOWN POLICE				
Calibration File No.:	01108	Calib. Date:	02/18/2015	Calib. No.:	00026
Certification File No.:	01109	Cert. Date:	02/18/2015	Cert. No.:	00021
Linearity File No.:	01110	Lin. Date:	02/18/2015	Lin. No.:	00021
Solution File No.:	01113	Soln. Date:	03/22/2015	Soln. No.:	00183
Sequential File No.:	01113	File Date:	03/22/2015		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0182
Control Solution %:	0.100%			Expires:	06/20/2016
Solution Control Lot:	14F130			Bottle No.:	0058

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:48S		
Control 1 EC	0.099%	02:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:49S		
Control 2 EC	0.098%	02:50S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:50S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:51S		
Control 3 EC	0.098%	02:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MILLER First Name: BENJAMIN MI: L

Signature: Badge No.: 1113

Date: 03/22/2015