

# Calibrating Unit New Standard Solution Report

|                         |                       |                         |
|-------------------------|-----------------------|-------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.: ARTN-0042   |
| Location:               | HIGHTSTOWN POLICE     |                         |
| Calibration File No.:   | 01108                 | Calib. Date: 02/18/2015 |
| Certification File No.: | 01109                 | Calib. No.: 00026       |
| Linearity File No.:     | 01110                 | Cert. Date: 02/18/2015  |
| Solution File No.:      | 01128                 | Cert. No.: 00021        |
| Sequential File No.:    | 01128                 | Lin. Date: 02/18/2015   |
|                         |                       | Lin. No.: 00021         |
|                         |                       | Soln. Date: 05/23/2015  |
|                         |                       | Soln. No.: 00186        |
|                         |                       | File Date: 05/23/2015   |

|                       |        |                  |                          |
|-----------------------|--------|------------------|--------------------------|
| Calibrating Unit:     | WET    | Model No.: CU-34 | Serial No.: DDUL S3-0182 |
| Control Solution %:   | 0.100% |                  | Expires: 06/20/2016      |
| Solution Control Lot: | 14F130 |                  | Bottle No.: 0068         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 22:19D |                |                     |
| Control 1 EC      | 0.098% | 22:20D | 34.1°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.099% | 22:20D | 34.1°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:21D |                |                     |
| Control 2 EC      | 0.098% | 22:21D | 34.1°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.099% | 22:21D | 34.1°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:22D |                |                     |
| Control 3 EC      | 0.098% | 22:23D | 34.1°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 22:23D | 34.1°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:23D |                |                     |

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature:  #1113

Badge No.: 1113

Date: 05/23/2015