Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTN-0042
Location:	HIGHTSTOWN POLICE				
Calibration File No.:	01134	Calib. Date:	07/13/2015	Calib. No.:	00027
Certification File No.:	01135	Cert. Date:	07/13/2015	Cert. No.:	00022
Linearity File No.:	01136	Lin. Date:	07/13/2015	Lin. No.:	00022
Solution File No.:	01146	Soln. Date:	09/10/2015	Soln. No.:	00190
Sequential File No.:	01146	File Date:	09/10/2015		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0182
Control Solution %:	0.100%			Expires:	06/20/2016
Solution Control Lot:	14F130			Bottle No.:	0067

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:45D		
Control 1 EC	0.101%	09:46D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:46D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:46D		
Control 2 EC	0.101%	09:47D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.102%	09:47D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:48D		
Control 3 EC	0.101%	09:48D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:48D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MILLER First Name: BENJAMIN MI: L

Signature: Sgt. J. Miller #1113 Badge No.: 1113

Date: 09/10/2015