

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01240	Calib. Date: 10/26/2016
Certification File No.:	01241	Calib. No.: 00030
Linearity File No.:	01242	Cert. Date: 10/26/2016
Solution File No.:	01247	Cert. No.: 00025
Sequential File No.:	01247	Lin. Date: 10/26/2016
		Lin. No.: 00025
		Soln. Date: 12/30/2016
		Soln. No.: 00213
		File Date: 12/30/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0182
Solution Control Lot:	201605D	Expires: 05/17/2018
		Bottle No.: 0796

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:19S		
Control 1 EC	0.101%	13:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:21S		
Control 2 EC	0.098%	13:22S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:22S		
Control 3 EC	0.099%	13:23S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: Sept 2016 #1113



Badge No.: 1113

Date: 12/30/2016