

# Calibrating Unit

## New Standard Solution Report

|                         |                       |                          |
|-------------------------|-----------------------|--------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.: ARTN-0042    |
| Location:               | HIGHTSTOWN POLICE     |                          |
| Calibration File No.:   | 01167                 | Calib. Date: 01/11/2016  |
| Certification File No.: | 01168                 | Calib. No.: 00028        |
| Linearity File No.:     | 01169                 | Cert. Date: 01/11/2016   |
| Solution File No.:      | 01199                 | Cert. No.: 00023         |
| Sequential File No.:    | 01199                 | Lin. Date: 01/11/2016    |
|                         |                       | Lin. No.: 00023          |
|                         |                       | Soln. Date: 04/09/2016   |
|                         |                       | Soln. No.: 00201         |
|                         |                       | File Date: 04/09/2016    |
| <br>                    |                       |                          |
| Calibrating Unit:       | WET                   | Model No.: CU-34         |
| Control Solution %:     | 0.100%                | Serial No.: DDUL S3-0183 |
| Solution Control Lot:   | 16A150                | Expires: 01/05/2017      |
|                         |                       | Bottle No.: 0503         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 17:09D |                |                     |
| Control 1 EC      | 0.101% | 17:10D | 33.9°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.102% | 17:10D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:11D |                |                     |
| Control 2 EC      | 0.100% | 17:12D | 33.9°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.101% | 17:12D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:12D |                |                     |
| Control 3 EC      | 0.101% | 17:13D | 33.9°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.101% | 17:13D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 17:14D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GORDY

First Name: MICHAEL

MI: A

Signature: \_\_\_\_\_

Badge No.: 1119

Date: 04/09/2016