

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01167	Calib. Date: 01/11/2016
Certification File No.:	01168	Calib. No.: 00028
Linearity File No.:	01169	Cert. Date: 01/11/2016
Solution File No.:	01203	Cert. No.: 00023
Sequential File No.:	01203	Lin. Date: 01/11/2016
		Lin. No.: 00023
		Soln. Date: 04/18/2016
		Soln. No.: 00203
		File Date: 04/18/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	15F139	Expires: 06/13/2017
		Bottle No.: 1367

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:19D		
Control 1 EC	0.099%	16:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:20D		
Control 2 EC	0.098%	16:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:22D		
Control 3 EC	0.098%	16:23D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	16:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GENDRON

First Name: FRANK

MI: G

Signature: \_\_\_\_\_

Badge No.: 1120

Date: 04/18/2016