



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lt. Governor

ROBERT T. LOUGY
Acting Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 04/13/2016

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 201603L

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1222 to 0.1232 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is March 22, 2018.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

[Handwritten signature of Ali M. Alaoui]

Ali M. Alaoui, Ph.D.
Research Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 14 day of April, 2016.

[Handwritten signature of John R. Leaver]
Notary
JOHN R LEAVER
ID # 2207138
NOTARY PUBLIC
STATE OF NEW JERSEY
My Commission Expires Dec. 14, 2017



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



Packing Slip

Customer no Order No Order date
 150054156 10848360 04/22/2016

Packing slip no. Ship date
 81025691 04/22/2016
 Please reference on inquiries

Ship to
 150054156
 BOROUGH OF HIGHTSTOWN POLICE DEPT.
 ATTN: SGT. MARK NIRO
 MERCER COUNTY
 415A MERCER STREET
 HIGHTSTOWN, NJ 08520
 USA

Payer

Your Purchase Order Number

15-01332
 Any questions? Please contact:
 ERIKA ESQUEDA

Customer
 BOROUGH OF HIGHTSTOWN POLICE DEPT.
 ATTN: SGT. MARK NIRO
 MERCER COUNTY
 415A MERCER STREET
 HIGHTSTOWN, NJ 08520
 USA

Your sales person

CHRISTOPHER FAIRCHILD
 Phone: 866-385-5900
 Fax: 972-929-1260

Delivery terms
 FA
 FREIGHT ALLOWED

| Item# | sh. Quant. | Part no. Description |
|-------|------------|--|
| | | Ship via: Fed Ex Ground |
| | | NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900. |
| | | ORIGINAL ORDER: 10798169. CUSTOMER RECEIVED BAD CASE OF SOLUTION, LOT # 16A150. SENDING 6 BOTTLES OF SOLUTION, BACKORDERING 6. ORIGINAL ORDER WAS FOR 15, NEW VENDOR CASE COMES IN 12, ISSUED A CREDIT FOR DIFFERENCE OF 3 BOTTLES UNDER S/O 60424671. |
| 0010 | 6 EA | 4407061 CERT SOLUTION, .10% NEW JERSEY CERTIFIED Ordered / Back ordered : 0 / 0 EA |

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 536410
 Pittsburgh, PA 15253-5906