

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0042
Location:	HIGHTSTOWN POLICE		
Calibration File No.:	01206	Calib. Date:	05/06/2016
Certification File No.:	01207	Cert. No.:	00029
Linearity File No.:	01208	Cert. Date:	05/06/2016
Solution File No.:	01237	Lin. Date:	05/06/2016
Sequential File No.:	01237	Soln. Date:	10/03/2016
		File Date:	10/03/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0183
Solution Control Lot:	201605D	Expires:	05/17/2018
		Bottle No.:	0793

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:33D		
Control 1 EC	0.101%	12:34D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:35D		
Control 2 EC	0.099%	12:35D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	12:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:36D		
Control 3 EC	0.100%	12:37D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	12:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: *Sgt. J Miller #1113*

Badge No.: 1113

Date: 10/03/2016