

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01167	Calib. Date: 01/11/2016
Certification File No.:	01168	Calib. No.: 00028
Linearity File No.:	01169	Cert. Date: 01/11/2016
Solution File No.:	01175	Cert. No.: 00023
Sequential File No.:	01175	Lin. Date: 01/11/2016
		Lin. No.: 00023
		Soln. Date: 02/01/2016
		Soln. No.: 00196
		File Date: 02/01/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	15F139	Expires: 06/13/2017
		Bottle No.: 1374

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:48S		
Control 1 EC	0.097%	09:49S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:49S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:50S		
Control 2 EC	0.096%	09:50S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:51S		
Control 3 EC	0.096%	09:52S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	09:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: \_\_\_\_\_



Badge No.: 1113

Date: 02/01/2016