

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0042  
Location: HIGHTSTOWN POLICE  
Calibration File No.: 01167 Calib. Date: 01/11/2016 Calib. No.: 00028  
Certification File No.: 01168 Cert. Date: 01/11/2016 Cert. No.: 00023  
Linearity File No.: 01169 Lin. Date: 01/11/2016 Lin. No.: 00023  
Solution File No.: 01177 Soln. Date: 02/02/2016 Soln. No.: 00197  
Sequential File No.: 01177 File Date: 02/02/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0183  
Control Solution %: 0.100% Expires: 06/13/2017  
Solution Control Lot: 15F139 Bottle No.: 1364

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:41S		
Control 1 EC	0.096%	07:42S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	07:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:43S		
Control 2 EC	0.095%	07:44S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:44S		
Control 3 EC	0.096%	07:45S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	07:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:46S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: 

Badge No.: 1113

Date: 02/02/2016