

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01167	Calib. Date: 01/11/2016
Certification File No.:	01168	Calib. No.: 00028
Linearity File No.:	01169	Cert. Date: 01/11/2016
Solution File No.:	01180	Cert. No.: 00023
Sequential File No.:	01180	Lin. Date: 01/11/2016
		Lin. No.: 00023
		Soln. Date: 02/09/2016
		Soln. No.: 00198
		File Date: 02/09/2016

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUL S3-0183
Control Solution %:	0.100%		Expires: 06/13/2017
Solution Control Lot:	15F139		Bottle No.: 1372

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:20S		
Control 1 EC	0.098%	13:21S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	13:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:22S		
Control 2 EC	0.096%	13:23S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:24S		
Control 3 EC	0.096%	13:24S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: GENDRON

First Name: FRANK

MI: G

Signature: _____



Badge No.: 1120

Date: 02/09/2016