

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0042
Location: HIGHTSTOWN POLICE
Calibration File No.: 01206 Calib. Date: 05/06/2016 Calib. No.: 00029
Certification File No.: 01207 Cert. Date: 05/06/2016 Cert. No.: 00024
Linearity File No.: 01208 Lin. Date: 05/06/2016 Lin. No.: 00024
Solution File No.: 01215 Soln. Date: 06/07/2016 Soln. No.: 00206
Sequential File No.: 01215 File Date: 06/07/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0183
Control Solution %: 0.100% Expires: 03/22/2018
Solution Control Lot: 201603L Bottle No.: 1448

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:45D		
Control 1 EC	0.102%	16:46D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:47D		
Control 2 EC	0.100%	16:47D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:48D		
Control 3 EC	0.100%	16:49D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: Sgt. J. Lewis 1113

Badge No.: 1113

Date: 06/07/2016