

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01318	Calib. Date: 01/08/2018
Certification File No.:	01319	Cert. No.: 00033
Linearity File No.:	01320	Cert. Date: 01/08/2018
Solution File No.:	01327	Lin. No.: 00028
Sequential File No.:	01327	Lin. Date: 01/08/2018
		Soln. No.: 00231
		Soln. Date: 01/24/2018
		File Date: 01/24/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	17290	Expires: 09/14/2019
		Bottle No.: 0842

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:42S		
Control 1 EC	0.101%	01:43S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	01:43S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:44S		
Control 2 EC	0.098%	01:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:46S		
Control 3 EC	0.099%	01:47S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	01:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: JIMENEZ

First Name: SGT.

MI: F.

Signature: \_\_\_\_\_



Badge No.: 1110

Date: 01/24/2018