

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01318	Calib. Date: 01/08/2018
Certification File No.:	01319	Calib. No.: 00033
Linearity File No.:	01320	Cert. Date: 01/08/2018
Solution File No.:	01353	Cert. No.: 00028
Sequential File No.:	01353	Lin. Date: 01/08/2018
		Lin. No.: 00028
		Soln. Date: 03/31/2018
		Soln. No.: 00234
		File Date: 03/31/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	17290	Expires: 09/14/2019
		Bottle No.: 0729

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:54S		
Control 1 EC	0.100%	20:54S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:55S		
Control 2 EC	0.098%	20:56S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	20:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:57S		
Control 3 EC	0.098%	20:58S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: \_\_\_\_\_

Badge No.: 1123

Date: 03/31/2018