

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01318	Calib. Date: 01/08/2018
Certification File No.:	01319	Cert. No.: 00033
Linearity File No.:	01320	Cert. Date: 01/08/2018
Solution File No.:	01374	Lin. Date: 01/08/2018
Sequential File No.:	01374	Soln. Date: 05/22/2018
		Lin. No.: 00028
		Soln. No.: 00237
		File Date: 05/22/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	17290	Expires: 09/14/2019
		Bottle No.: 0658

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:57D		
Control 1 EC	0.101%	12:58D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:59D		
Control 2 EC	0.100%	12:59D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:00D		
Control 3 EC	0.100%	13:01D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:02D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: 

Badge No.: 1113

Date: 05/22/2018