

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0042
Location:	HIGHTSTOWN POLICE		
Calibration File No.:	01477	Calib. Date:	10/04/2019
Certification File No.:	01478	Cert. No.:	00037
Linearity File No.:	01479	Cert. Date:	10/04/2019
Solution File No.:	01505	Lin. No.:	00032
Sequential File No.:	01505	Soln. Date:	01/23/2020
		Soln. No.:	00264
		File Date:	01/23/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0183
Solution Control Lot:	19300	Expires:	10/30/2021
		Bottle No.:	0563

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:25S		
Control 1 EC	0.100%	22:26S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	22:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:27S		
Control 2 EC	0.098%	22:27S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:28S		
Control 3 EC	0.098%	22:29S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN	First Name: SGT	MI: R
Signature: 	Badge No.: 1123	Date: 01/23/2020