

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0042
Location:	HIGHTSTOWN POLICE		
Calibration File No.:	01524	Calib. Date:	07/23/2020
Certification File No.:	01525	Cert. No.:	00039
Linearity File No.:	01526	Cert. Date:	07/23/2020
Solution File No.:	01541	Lin. No.:	00034
Sequential File No.:	01541	Lin. Date:	07/23/2020
		Soln. Date:	10/20/2020
		Soln. No.:	00276
		File Date:	10/20/2020

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0183
Control Solution %:	0.100%			Expires:	07/06/2022
Solution Control Lot:	20320			Bottle No.:	1362

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:27D		
Control 1 EC	0.099%	10:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:29D		
Control 2 EC	0.099%	10:29D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:30D		
Control 3 EC	0.099%	10:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: _____

Syl J. Miller #1113

Badge No.: 1113

Date: 10/20/2020