

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0042
Location: HIGHTSTOWN POLICE
Calibration File No.: 01477 Calib. Date: 10/04/2019 Calib. No.: 00037
Certification File No.: 01478 Cert. Date: 10/04/2019 Cert. No.: 00032
Linearity File No.: 01479 Lin. Date: 10/04/2019 Lin. No.: 00032
Solution File No.: 01507 Soln. Date: 02/23/2020 Soln. No.: 00265
Sequential File No.: 01507 File Date: 02/23/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0183
Control Solution %: 0.100% Expires: 10/30/2021
Solution Control Lot: 19300 Bottle No.: 0560

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:57S		
Control 1 EC	0.101%	04:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:59S		
Control 2 EC	0.099%	04:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:00S		
Control 3 EC	0.100%	05:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:02S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JIMENEZ

First Name: SGT.

MI: F.

Signature: 

Badge No.: 1110

Date: 02/23/2020