

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01508	Calib. Date: 03/02/2020
Certification File No.:	01509	Calib. No.: 00038
Linearity File No.:	01510	Cert. Date: 03/02/2020
Solution File No.:	01514	Cert. No.: 00033
Sequential File No.:	01514	Lin. Date: 03/02/2020
		Lin. No.: 00033
		Soln. Date: 03/16/2020
		Soln. No.: 00267
		File Date: 03/16/2020

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0182
Control Solution %:	0.100%			Expires:	10/30/2021
Solution Control Lot:	19300			Bottle No.:	0499

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:03S		
Control 1 EC	0.097%	03:04S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:05S		
Control 2 EC	0.098%	03:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:07S		
Control 3 EC	0.098%	03:07S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: \_\_\_\_\_

Badge No.: 1123

Date: 03/16/2020