

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01508	Calib. Date: 03/02/2020
Certification File No.:	01509	Calib. No.: 00038
Linearity File No.:	01510	Cert. Date: 03/02/2020
Solution File No.:	01515	Cert. No.: 00033
Sequential File No.:	01515	Lin. Date: 03/02/2020
		Lin. No.: 00033
		Soln. Date: 04/14/2020
		Soln. No.: 00268
		File Date: 04/14/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0182
Solution Control Lot:	19300	Expires: 10/30/2021
		Bottle No.: 0411

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:25D		
Control 1 EC	0.101%	00:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:26D		
Control 2 EC	0.100%	00:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:28D		
Control 3 EC	0.099%	00:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JIMENEZ

First Name: SGT.

MI: F.

Signature: _____

Badge No.: 1110

Date: 04/14/2020