

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0042  
Location: HIGHTSTOWN POLICE  
Calibration File No.: 01508 Calib. Date: 03/02/2020 Calib. No.: 00038  
Certification File No.: 01509 Cert. Date: 03/02/2020 Cert. No.: 00033  
Linearity File No.: 01510 Lin. Date: 03/02/2020 Lin. No.: 00033  
Solution File No.: 01517 Soln. Date: 05/11/2020 Soln. No.: 00269  
Sequential File No.: 01517 File Date: 05/11/2020

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0182  
Control Solution %: 0.100% Expires: 10/30/2021  
Solution Control Lot: 19300 Bottle No.: 0507

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:15D		
Control 1 EC	0.101%	04:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:16D		
Control 2 EC	0.100%	04:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:18D		
Control 3 EC	0.100%	04:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: \_\_\_\_\_



Badge No.: 1123

Date: 05/11/2020