

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARTN-0042 |
| Location: | HIGHTSTOWN POLICE | |
| Calibration File No.: | 01508 | Calib. Date: 03/02/2020 |
| Certification File No.: | 01509 | Cert. No.: 00038 |
| Linearity File No.: | 01510 | Cert. Date: 03/02/2020 |
| Solution File No.: | 01519 | Lin. Date: 03/02/2020 |
| Sequential File No.: | 01519 | Lin. No.: 00033 |
| | | Soln. Date: 06/08/2020 |
| | | Soln. No.: 00270 |
| | | File Date: 06/08/2020 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUL S3-0182 |
| Solution Control Lot: | 19300 | Expires: 10/30/2021 |
| | | Bottle No.: 0559 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 04:35D | | |
| Control 1 EC | 0.100% | 04:36D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 04:36D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:37D | | |
| Control 2 EC | 0.099% | 04:38D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 04:38D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:38D | | |
| Control 3 EC | 0.100% | 04:39D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 04:39D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:40D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: _____

Badge No.: 1123

Date: 06/08/2020