

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|-----------------------|--------------|------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARTN-0042 |
| Location: | HIGHTSTOWN POLICE | | |
| Calibration File No.: | 01524 | Calib. Date: | 07/23/2020 |
| Certification File No.: | 01525 | Cert. Date: | 07/23/2020 |
| Linearity File No.: | 01526 | Lin. Date: | 07/23/2020 |
| Solution File No.: | 01535 | Soln. Date: | 08/22/2020 |
| Sequential File No.: | 01535 | File Date: | 08/22/2020 |

| | | | | | |
|-----------------------|--------|------------|-------|-------------|--------------|
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDUL S3-0183 |
| Control Solution %: | 0.100% | | | Expires: | 07/06/2022 |
| Solution Control Lot: | 20320 | | | Bottle No.: | 1352 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 00:39D | | |
| Control 1 EC | 0.100% | 00:39D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 00:39D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 00:40D | | |
| Control 2 EC | 0.099% | 00:41D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 00:41D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 00:42D | | |
| Control 3 EC | 0.099% | 00:42D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 00:42D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 00:43D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JIMENEZ

First Name: SGT.

MI: F.

Signature: _____

SGT. F. Jimenez

1110

Badge No.: 1110

Date: 08/22/2020

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARTN-0042 |
| Location: | HIGHTSTOWN POLICE | |
| Calibration File No.: | 01524 | Calib. Date: 07/23/2020 |
| Certification File No.: | 01525 | Calib. No.: 00039 |
| Linearity File No.: | 01526 | Cert. Date: 07/23/2020 |
| Solution File No.: | 01534 | Cert. No.: 00034 |
| Sequential File No.: | 01534 | Lin. Date: 07/23/2020 |
| | | Lin. No.: 00034 |
| | | Soln. Date: 08/21/2020 |
| | | Soln. No.: 00273 |
| | | File Date: 08/21/2020 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUL S3-0183 |
| Solution Control Lot: | 20320 | Expires: 07/06/2022 |
| | | Bottle No.: 1344 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|-----------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 21:09D | | |
| Control 1 EC | ----% | 21:09D | 34.0°C | CTRL GAS SUPPLY |
| Control 1 IR | ----% | 21:09D | 34.0°C | CTRL GAS SUPPLY |
| Ambient Air Blank | ----% | --:-- | | |
| Control 2 EC | ----% | --:-- | N/A | |
| Control 2 IR | ----% | --:-- | N/A | |
| Ambient Air Blank | ----% | --:-- | | |
| Control 3 EC | ----% | --:-- | N/A | |
| Control 3 IR | ----% | --:-- | N/A | |
| Ambient Air Blank | ----% | --:-- | | |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JIMENEZ

First Name: SGT.

MI: F.

Signature: _____

Badge No.: 1110

Date: 08/21/2020