

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0042
Location:	HIGHTSTOWN POLICE		
Calibration File No.:	01524	Calib. Date:	07/23/2020
Certification File No.:	01525	Cert. Date:	07/23/2020
Linearity File No.:	01526	Lin. Date:	07/23/2020
Solution File No.:	01538	Soln. Date:	09/21/2020
Sequential File No.:	01538	File Date:	09/21/2020

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0183
Control Solution %:	0.100%			Expires:	07/06/2022
Solution Control Lot:	20320			Bottle No.:	1341

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:47D		
Control 1 EC	0.100%	03:48D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:49D		
Control 2 EC	0.097%	03:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:50D		
Control 3 EC	0.098%	03:51D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: JIMENEZ

First Name: SGT.

MI: F.

Signature: _____



Badge No.: 1110

Date: 09/21/2020