Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0042
Location: HIGHTSTOWN POLICE

Calibration File No.:	01586	Calib. Date:	06/17/2021	Calib. No.:	00041
Certification File No.:	01587	Cert. Date:	06/17/2021	Cert. No.:	00036
Linearity File No.:	01588	Lin. Date:	06/17/2021	Lin. No.:	00036
Solution File No.:	01600	Soln. Date:	07/24/2021	Soln. No.:	00287
Sequential File No.:	01600	File Date:	07/24/2021		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0182
Control Solution %:	0.100%			Expires:	05/04/2023
Solution Control Lot:	21150			Bottle No.:	1308

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:01D		
Control 1 EC	0.100%	05:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	05:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:03D		
Control 2 EC	0.099%	05:03D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	05:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:04D		
Control 3 EC	0.098%	05:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: BUCK First Name: PTL. MI: R

Signature: Ptl. TB #1129 Badge No.: 1129
Date: 07/24/2021