

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01550	Calib. Date: 01/07/2021
Certification File No.:	01551	Calib. No.: 00040
Linearity File No.:	01552	Cert. Date: 01/07/2021
Solution File No.:	01553	Cert. No.: 00035
Sequential File No.:	01553	Lin. Date: 01/07/2021
		Lin. No.: 00035
		Soln. Date: 01/07/2021
		Soln. No.: 00279
		File Date: 01/07/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	20320	Expires: 07/06/2022
		Bottle No.: 1322

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:08S		
Control 1 EC	0.101%	11:08S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:09S		
Control 2 EC	0.101%	11:10S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:11S		
Control 3 EC	0.100%	11:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJSP2-055 MRW

**Changed By:**

Last Name: WATSON

First Name: MATTHEW

MI: R

Signature:  #2078

Badge No.: 7078

Date: 01/07/2021