

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0042
Location: HIGHTSTOWN POLICE
Calibration File No.: 01586 Calib. Date: 06/17/2021 Calib. No.: 00041
Certification File No.: 01587 Cert. Date: 06/17/2021 Cert. No.: 00036
Linearity File No.: 01588 Lin. Date: 06/17/2021 Lin. No.: 00036
Solution File No.: 01611 Soln. Date: 11/20/2021 Soln. No.: 00291
Sequential File No.: 01611 File Date: 11/20/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0182
Control Solution %: 0.100% Expires: 05/04/2023
Solution Control Lot: 21150 Bottle No.: 0784

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:47S		
Control 1 EC	0.100%	13:47S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48S		
Control 2 EC	0.100%	13:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:50S		
Control 3 EC	0.100%	13:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: _____

 #1123

Badge No.: 1123

Date: 11/20/2021