

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01550	Calib. Date: 01/07/2021
Certification File No.:	01551	Cert. No.: 00040
Linearity File No.:	01552	Cert. Date: 01/07/2021
Solution File No.:	01558	Lin. No.: 00035
Sequential File No.:	01558	Lin. Date: 01/07/2021
		Soln. No.: 00281
		Soln. Date: 03/05/2021
		File Date: 03/05/2021

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	20320	Expires: 07/06/2022
		Bottle No.: 1382

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:16S		
Control 1 EC	0.102%	04:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	04:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:18S		
Control 2 EC	0.099%	04:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	04:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:20S		
Control 3 EC	0.099%	04:21S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	04:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:22S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: \_\_\_\_\_

Badge No.: 1123

Date: 03/05/2021