

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0042
Location:	HIGHTSTOWN POLICE	Calib. Date:	01/07/2021
Calibration File No.:	01550	Calib. No.:	00040
Certification File No.:	01551	Cert. Date:	01/07/2021
Linearity File No.:	01552	Lin. Date:	01/07/2021
Solution File No.:	01563	Soln. Date:	04/02/2021
Sequential File No.:	01563	File Date:	04/02/2021

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0183
Control Solution %:	0.100%			Expires:	07/06/2022
Solution Control Lot:	20320			Bottle No.:	1325

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:45S		
Control 1 EC	0.103%	18:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:47S		
Control 2 EC	0.100%	18:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	18:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:48S		
Control 3 EC	0.101%	18:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.104%	18:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: JIMENEZ

First Name: SGT.

MI: F.

Signature:

Badge No.: 1110

Date: 04/02/2021