

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0042
Location:	HIGHTSTOWN POLICE		
Calibration File No.:	01550	Calib. Date:	01/07/2021
Certification File No.:	01551	Calib. No.:	00040
Linearity File No.:	01552	Cert. Date:	01/07/2021
Solution File No.:	01567	Cert. No.:	00035
Sequential File No.:	01567	Lin. Date:	01/07/2021
		Lin. No.:	00035
		Soln. Date:	04/30/2021
		Soln. No.:	00283
		File Date:	04/30/2021
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0183
Solution Control Lot:	20320	Expires:	07/06/2022
		Bottle No.:	1397

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:48D		
Control 1 EC	0.100%	19:49D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:50D		
Control 2 EC	0.100%	19:51D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:52D		
Control 3 EC	0.101%	19:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	19:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JIMENEZ

First Name: SGT.

MI: F.

Signature:

Badge No.: 1110

Date: 04/30/2021