Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTN-0042
Location:	HIGHTSTOWN POLICE				
Calibration File No.:	01586	Calib. Date:	06/17/2021	Calib. No.:	00041
Certification File No.:	01587	Cert. Date:	06/17/2021	Cert. No.:	00036
Linearity File No.:	01588	Lin. Date:	06/17/2021	Lin. No.:	00036
Solution File No.:	01589	Soln. Date:	06/17/2021	Soln. No.:	00285
Sequential File No.:	01589	File Date:	06/17/2021		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0182
Control Solution %:	0.100%			Expires:	05/04/2023
Solution Control Lot:	21150			Bottle No.:	1436

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:42D		
Control 1 EC	0.100%	14:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:43D		
Control 2 EC	0.099%	14:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:45D		
Control 3 EC	0.100%	14:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJP2-035 (MRW)

Changed By:

Last Name: WATSON First Name: MATTHEW MI: R

Signature: Tpr I #7078 Badge No.: 7078

Date: 06/17/2021