

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01586	Calib. Date: 06/17/2021
Certification File No.:	01587	Calib. No.: 00041
Linearity File No.:	01588	Cert. Date: 06/17/2021
Solution File No.:	01605	Cert. No.: 00036
Sequential File No.:	01605	Lin. Date: 06/17/2021
		Lin. No.: 00036
		Soln. Date: 09/21/2021
		Soln. No.: 00289
		File Date: 09/21/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0182
Solution Control Lot:	21150	Expires: 05/04/2023
		Bottle No.: 1426

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:33D		
Control 1 EC	0.100%	15:34D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:34D		
Control 2 EC	0.099%	15:35D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:36D		
Control 3 EC	0.099%	15:37D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: _____

Badge No.: 1123

Date: 09/21/2021