

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0042
Location: HIGHTSTOWN POLICE
Calibration File No.: 01620 Calib. Date: 12/21/2021 Calib. No.: 00043
Certification File No.: 01621 Cert. Date: 12/21/2021 Cert. No.: 00038
Linearity File No.: 01622 Lin. Date: 12/21/2021 Lin. No.: 00038
Solution File No.: 01627 Soln. Date: 01/17/2022 Soln. No.: 00294
Sequential File No.: 01627 File Date: 01/17/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0183
Control Solution %: 0.100% Expires: 05/04/2023
Solution Control Lot: 21150 Bottle No.: 1315

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:25S		
Control 1 EC	0.098%	07:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:27S		
Control 2 EC	0.095%	07:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:28S		
Control 3 EC	0.096%	07:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: _____



Badge No.: 1123

Date: 01/17/2022