

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0042  
Location: HIGHTSTOWN POLICE  
Calibration File No.: 01673 Calib. Date: 09/02/2022 Calib. No.: 00045  
Certification File No.: 01674 Cert. Date: 09/02/2022 Cert. No.: 00040  
Linearity File No.: 01675 Lin. Date: 09/02/2022 Lin. No.: 00040  
Solution File No.: 01678 Soln. Date: 10/05/2022 Soln. No.: 00309  
Sequential File No.: 01678 File Date: 10/05/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0182  
Control Solution %: 0.100% Expires: 05/09/2024  
Solution Control Lot: 22160 Bottle No.: 1257

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:42D		
Control 1 EC	0.101%	07:42D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:43D		
Control 2 EC	0.099%	07:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:45D		
Control 3 EC	0.101%	07:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature:  #1113

Badge No.: 1113

Date: 10/05/2022