Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTN-0042
Location:	HIGHTSTOWN POLICE				
Calibration File No.:	01673	Calib. Date:	09/02/2022	Calib. No.:	00045
Certification File No.:	01674	Cert. Date:	09/02/2022	Cert. No.:	00040
Linearity File No.:	01675	Lin. Date:	09/02/2022	Lin. No.:	00040
Solution File No.:	01680	Soln. Date:	11/02/2022	Soln. No.:	00310
Sequential File No.:	01680	File Date:	11/02/2022		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0182
Control Solution %:	0.100%			Expires:	05/09/2024
Solution Control Lot:	22160			Bottle No.:	0927

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:22S		
Control 1 EC	0.100%	17:23S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	17:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:24S		
Control 2 EC	0.100%	17:25S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	17:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:25S		
Control 3 EC	0.100%	17:26S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	17:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:27S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN First Name: SGT MI: R

Badge No.: 1123

Signature: #1123 Date: 11/02/2022