

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0042
Location:	HIGHTSTOWN POLICE		
Calibration File No.:	01620	Calib. Date:	12/21/2021
Certification File No.:	01621	Cert. Date:	12/21/2021
Linearity File No.:	01622	Lin. Date:	12/21/2021
Solution File No.:	01634	Soln. Date:	03/12/2022
Sequential File No.:	01634	File Date:	03/12/2022
		Calib. No.:	00043
		Cert. No.:	00038
		Lin. No.:	00038
		Soln. No.:	00297
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0183
Solution Control Lot:	21480	Expires:	12/28/2023
		Bottle No.:	1366

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:47S		
Control 1 EC	-.---%	06:48S	N/A	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	06:48S	N/A	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: _____

Badge No.: 1113

Date: 03/12/2022