

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01620	Calib. Date: 12/21/2021
Certification File No.:	01621	Calib. No.: 00043
Linearity File No.:	01622	Cert. Date: 12/21/2021
Solution File No.:	01636	Cert. No.: 00038
Sequential File No.:	01636	Lin. Date: 12/21/2021
		Lin. No.: 00038
		Soln. Date: 03/12/2022
		Soln. No.: 00299
		File Date: 03/12/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	21480	Expires: 12/28/2023
		Bottle No.: 1326

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:20S		
Control 1 EC	0.101%	08:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:21S		
Control 2 EC	0.098%	08:22S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:23S		
Control 3 EC	0.098%	08:24S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: _____



Badge No.: 1113

Date: 03/12/2022