

# Calibrating Unit

## New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01620	Calib. Date: 12/21/2021
Certification File No.:	01621	Calib. No.: 00043
Linearity File No.:	01622	Cert. Date: 12/21/2021
Solution File No.:	01640	Cert. No.: 00038
Sequential File No.:	01640	Lin. Date: 12/21/2021
		Lin. No.: 00038
		Soln. Date: 04/02/2022
		Soln. No.: 00300
		File Date: 04/02/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	21480	Expires: 12/28/2023
		Bottle No.: 1333

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:31S		
Control 1 EC	0.100%	04:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:33S		
Control 2 EC	0.098%	04:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:34S		
Control 3 EC	0.098%	04:35S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: STEPHENSEN	First Name: SGT	MI: R
Signature: 	# 1123	Badge No.: 1123
		Date: 04/02/2022