

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01648	Calib. Date: 05/24/2022
Certification File No.:	01649	Calib. No.: 00044
Linearity File No.:	01650	Cert. Date: 05/24/2022
Solution File No.:	01666	Cert. No.: 00039
Sequential File No.:	01666	Lin. Date: 05/24/2022
		Lin. No.: 00039
		Soln. Date: 07/11/2022
		Soln. No.: 00306
		File Date: 07/11/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	22070	Expires: 02/17/2024
		Bottle No.: 0589

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:57D		
Control 1 EC	0.100%	00:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:59D		
Control 2 EC	0.099%	00:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:00D		
Control 3 EC	0.099%	01:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: \_\_\_\_\_

Badge No.: 1123

Date: 07/11/2022