

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01648	Calib. Date: 05/24/2022
Certification File No.:	01649	Calib. No.: 00044
Linearity File No.:	01650	Cert. Date: 05/24/2022
Solution File No.:	01672	Cert. No.: 00039
Sequential File No.:	01672	Lin. Date: 05/24/2022
		Lin. No.: 00039
		Soln. Date: 08/08/2022
		Soln. No.: 00307
		File Date: 08/08/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	22070	Expires: 02/17/2024
		Bottle No.: 0590

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:46D		
Control 1 EC	0.099%	02:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:48D		
Control 2 EC	0.098%	02:49D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:49D		
Control 3 EC	0.098%	02:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN First Name: SGT MI: R

Signature: _____

Badge No.: 1123
Date: 08/08/2022