

State of New Jersey Office of the Attorney General

ÖFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068

JEFFREY S. CHIESA
Attorney General

COLONEL JOSEPH R. FUENTES

Superintendent

KIM GUADAGNO

Lt. Governor

CHRIS CHRISTIE

Governor

CERTIFICATION OF ANALYSIS 0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

(609) 882-2000

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 10/3/2012

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 121107

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of <u>0.1211</u> to <u>0.1218</u> grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is <u>September 13, 2014</u>.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaouie, Ph.D. Research Scientist

NJSP Office of Forensic Sciences

Sworm to and subscribed before me this 5 day of October, 2012.

Notary

Undo L Deserts Matery Public, New Jersey My Development Bydras & 17-14

(PA)

"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer Printed on Recycled Paper and Recycloble



Packing Slip

Customer no Order No 150054156 10597317 Order date

01/08/2013

Ship date Packing slip no. 01/08/2013 80742581

Please reference on inquiries

ship to

150054156

BOROUGH OF HIGHTSTOWN POLICE DEPT. ATTN: SGT. MARK NIRO MERCER COUNTY 415A MERCER STREET HIGHTSTOWN, NJ 08520

USA

Your Purchase Order Number

01/08/2013 NANCY KISLAN

Any questions? Please contact:

MARA HERRERA

Your sales person

Phone: 973-398-3228

Customer

bill to

BOROUGH OF HIGHTSTOWN POLICE DEPT. ATTN: SGT. MARK NIRO MERCER COUNTY

415A MERCER STREET HIGHTSTOWN, NJ 08520

USA

Don Pouliot

Delivery terms

FA

FREIGHT ALLOWED

Page 1 / 1

Item#	sh. Quant.	Ship via: Fed Ex Ground NOTE: All claims for shortages/defects must be reported within 10 Business		
	:	days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.		
0010	25 EA	4407061 CERT SOLUTION, 10% NEW JERSEY CERTIFIED Ordered / Back ordered : 25 / 0 EA		
0020	1 EA	FREIGHT Freight Charges Ordered / Back ordered : 1 / 0 EA		

Draeger Safety Diagnostics Inc. Accounting Address: 101 Technology Drive Pittsburgh, PA 15275 Tel: (412) 788-5537 Fax: (412) 788-5598

Remit Wire Transfers To: Citizens Bank Acct. # 6209426615

Acct. Name: Draeger Safety Diagnostics Inc. Transit Number: 036076150

Federal ID Number: 84-1600159

REMIT TO: Draeger Safety Diagnostics Inc. P.O. Box 200337 Pittsburgh, PA 15251-0337